

# **RESOURCE Q column**

# **Product Information**

Cat#No# RE-272P

#### **Product Overview**

RESOURCE Q columns are prepacked with SOURCE 15Q, a strong anion exchanger for high resolution polishing purification of proteins at high flow rates.

#### Characteristic

Quaternary ammonium (Q) strong anion exchanger for high resolution ion exchange chromatography. Mono-sized 15 µm rigid beads give low back pressure and high flow rates allowing for high productivity. High chemical stability enables wide range of working conditions with good resistance to cleaning conditions at high pH.

## **Maximum operating pressure**

15 bar [1.5 MPa] (218 psi)

#### **Matrix**

Spherical and monodisperse, porous, rigid, polystyrene/divinyl benzene particles.

## Average particle size

~ 15 µm

## Ligand

Quaternary ammonium (strong anion exchanger)

## **Dynamic binding capacity**

40-54 mg BSA/ml

#### Recommended flow rate

<1800 cm/h

## Recommended column height

30 mm

Fax:1-631-938-8127 45-1 Ramsey Road, Shirley, NY 11967, USA



# **RESOURCE Q column**

## **Chemical stability**

Stable to commonly used aqueous buffers, 1.0 M HCl, 100% ethanol, 100% isopropanol, 100% acetonitril, 100% methanol, 1.0 M NaOH.

## pH working range

2 0 12

#### **CIP** stability

1 to 14

## **Temperature stability**

4°C to 40°C

#### **Storage**

4 to 30°C, 20% Ethanol

#### Cleaning-in-place

- 1. Wash with at least 2 column volumes of 2 M NaCl at 1 mL/min (RESOURCE 1 mL), 6 mL/min (RESOURCE 6 mL).
- 2. Wash with at least 4 column volumes of 1.0 M NaOH (same flow as in step 1).
- 3. Wash with at least 2 column volumes of 2 M NaCl (same flow as in step 1).
- 4. Rinse with at least 2 column volumes of distilled water (same flow as in step 1) until the UV-baseline and the eluent pH are stable.
- 5. Wash with at least 4 column volumes of start buffer or storage buffer (same flow as in step 1) until eluent pH and conductivity have reached the required values.

#### Pack size

1 mL

## Maximum flow velocity

10 mL/min

#### **Dimensions**

 $6.4 \times 30 \text{ mm}$ 

Fax:1-631-938-8127 45-1 Ramsey Road, Shirley, NY 11967, USA



# **RESOURCE Q column**

Column i.d.

6.4 mm

Fax:1-631-938-8127 45-1 Ramsey Road, Shirley, NY 11967, USA