

## HiScreen SP FF

### Product Information

---

**Cat#No#** Hi-223P

---

### Product Overview

---

HiScreen SP FF columns are prepacked with SP Sepharose Fast Flow, a strong cation exchanger, and are part of the process development platform available. The columns are an excellent choice for method optimization and parameter screening.

---

### Description

---

The small column volume of 4.7 mL and bed height of 10 cm make HiScreen columns excellent tools for method optimization, parameter screening, robustness testing, and convenient scale-up. Process fluid velocities can be applied, since the 10 cm bed height gives enough residence time and the results can then serve as basis for linear process scale-up.

---

### Characteristic

---

SP Sepharose Fast Flow is a strong cation exchanger and IEX Sepharose Fast Flow resin are well-proven, industrial standards for ion exchange chromatography.

Prepacked, ready-to-use columns for convenient process development optimization.

10 cm bed height of HiScreen columns is designed to allow method optimization and parameter screening.

Easily connected in series to achieve 20 cm bed height.

Small bed volume gives fast results and minimal sample/buffer consumption.

Reproducible results, scalable to BioProcess columns packed with the same resin using the same linear fluid velocity.

---

### Maximum operating pressure

---

1.5 bar [0.15 MPa] (22 psi)

---

### Sample preparation

---

1. Adjust the sample to the composition of the start buffer, using one of these methods: Dilute the sample with start buffer. Exchange buffer using a HiPrep 26/10 Desalting, HiTrap Desalting or PD-10 Desalting column.
2. Filter the sample through a 0.45 µm filter or centrifuge at 10 000 × g for 10 min immediately before loading it to the column. This prevents clogging and increases the life time of the column when loading large sample

## HiScreen SP FF

volumes.

---

**Metal ion capacity**

0.18 to 0.25 mmol H<sup>+</sup> /mL resin

---

**Matrix**

Cross-linked agarose, 6%, spherical

---

**Ionic Exchanger Type**

Strong cation

---

**Average particle size**

~ 90 µm

---

**Dynamic binding capacity**

~ 70 mg Ribonuclease A/mL resin

---

**Recommended flow rate**

< 450 cm/h

---

**Recommended column height**

100 mm

---

**Chemical stability**

Stable to commonly used aqueous buffers, 1 M NaOH, 8 M urea, 6 M guanidine hydrochloride, 70% ethanol.

---

**pH working range**

4 to 13

---

**CIP stability**

3 to 14

---

**Storage**

4 to 30°C, 20% Ethanol, 0.2 M Sodium Acetate.

---

**Elution buffer**

---

## HiScreen SP FF

50 mM sodium acetate, 1 M NaCl, pH 5.0.

---

### Cleaning-in-place

---

1. Wash with 2 column volumes (CV) 2 M NaCl. If detergents have been used, wash with 5 CV distilled water followed by 2 CV of 2 M NaCl.
  2. Re-equilibrate the column with at least 5 CV start buffer or until the UV baseline and pH/conductivity values are stable.
- 

### Purification procedures

---

1. Remove the stoppers and connect the column to the system. Avoid introducing air into the column.
  2. Wash with 1 column volume (CV) of distilled water. This step removes the ethanol and avoids the precipitation of buffer salts upon exposure to ethanol. The step can be omitted if precipitation is not likely to be a problem.
  3. Adjust the sample to the chosen starting pH and conductivity and load on the column.
  4. Wash with 5 to 10 CV start buffer or until the UV trace of the effluent returns to near baseline.
  5. Linear gradient elution Elute with 0 to 100% elution buffer (up to 1 M NaCl) in 10 to 20 CV. Step elution Elute with 5 CV elution buffer including NaCl at chosen concentration. Repeat at higher NaCl concentrations until the target protein has been eluted.
  6. Wash with 5 CV 1 M NaCl (100% elution buffer) to elute any remaining ionically bound material.
  7. If required, perform a CIP to clean the column.
  8. Re-equilibrate with 5 to 10 CV start buffer or until the UV baseline, eluent pH, and conductivity reach the required values.
- 

### Pack size

---

1 × 4.7 mL

---

### Maximum flow velocity

---

450 cm/h (3.5 mL/min)

---

### Dimensions

---

7.7 × 100 mm

---

### Column volume

---

## HiScreen SP FF

4.7 mL

---

**Column i.d.**

---

7.7 mm

---

**Column hardware pressure limit**

---

0.8 MPa (8 bar, 116 psi)

---

**Functional group**

---

-CH<sub>2</sub>CH<sub>2</sub>CH<sub>2</sub>SO<sub>3</sub> -

---