

QuantiChrom™ Copper Assay Kit (DICU-250)

Quantitative Colorimetric Copper Determination at 359nm

DESCRIPTION

Copper is an essential trace element. Copper-containing enzymes play important roles in iron and catecholamine metabolism, free radical scavenging, and in the synthesis of hemoglobin, elastin and collagen. Copper is mainly present in caeruloplasmin in the liver. Low levels of copper have been associated with mental retardation, depigmentation, anemia, hypotonia and scorbutic changes in bone. Levels of copper are key diagnostic indicator of diseases such as Wilson's disease, microcytic hypochromic anemia and bone disease due to reduced collagen synthesis.

Simple, direct and automation-ready procedures for measuring copper concentrations find wide applications in research, drug discovery and environmental monitoring. BioAssay Systems' copper assay kit is designed to measure copper with no or minimal sample treatment. The improved method utilizes a chromogen that forms a colored complex specifically with copper ions. The intensity of the color, measured at 359nm, is directly proportional to copper concentration in the sample. The optimized formulation substantially reduces interference by substances in the raw samples.

KEY FEATURES

Sensitive and accurate. Linear detection range 7 µg/dL (1.0 µM) to 300 µg/dL (47 µM) copper in 96-well plate assay.

Simple and high-throughput. The simple procedure can be readily automated as a high-throughput assay in 96-well plates for thousands of samples per day.

Improved reagent stability and versatility. The optimized formulation has greatly enhanced reagent and signal stability. Cuvet or 96-well plate assay.

APPLICATIONS

Direct Assays: biological, environmental, food and beverage samples.

Drug Discovery/Pharmacology: effects of drugs on Cu metabolism.

KIT CONTENTS (250 tests in 96-well plates)

Reagent A: 10 mL Reagent B: 1.5 mL Reagent C: 40 mL
Copper Standard: 1 mL 1.5 mg/dL Cu²⁺

Storage conditions. The kit is shipped at room temperature. Store all reagents at 4 °C. Shelf life: 6 months after receipt.

Precautions: reagents are for research use only. Normal precautions for laboratory reagents should be exercised while using the reagents. Please refer to Material Safety Data Sheet for detailed information.

PROCEDURES

Note: metal chelators (e.g. EDTA) interfere with this assay and should be avoided in sample preparation.

Procedure using 96-well plate:

1. Standards: transfer 100 µL dH₂O into one Eppendorf tube labeled "Blank". Into another tube labeled "Standard", mix 20 µL 1.5 mg/dL Standard and 80 µL dH₂O (final 300 µg/dL Cu²⁺).

Samples: transfer 100 µL samples into separate tubes.

Add 35 µL Reagent A (trichloroacetic acid) to each tube and mix by vortexing. If samples contain protein (e.g. serum/plasma), precipitates form. Centrifuge tubes for 2 min at 14,000 rpm and use clear supernatant for assay. For samples that do not contain protein, the mixture remains clear and centrifugation is not necessary.

Transfer 100 µL Blank, Standard and Sample into separate wells of a clear flat-bottom 96-well plate.

2. For each assay well, prepare Working Reagent by mixing 5 µL Reagent B and 150 µL Reagent C. Transfer 150 µL Working Reagent to each well and tap plate to mix thoroughly.

3. Incubate 5 min at room temperature and read optical density at 356-362nm (peak absorbance at 359nm).

Note: if sample OD values are higher than the OD value for the 300µg/dL Standard, dilute sample in dH₂O and repeat assay. Multiply the results by the dilution factor.

Procedure using cuvette:

Prepare standards and samples as for 96-well assay procedure.

1. Transfer 400 µL Standards and Samples into separate cuvetts.

2. Add 600 µL Working Reagent. Mix by pipetting.

3. Incubate 5 min at room temperature and read optical density at 356-362nm (peak absorbance at 359nm).

CALCULATION

The copper concentration of Sample is calculated as

$$= \frac{OD_{\text{SAMPLE}} - OD_{\text{BLANK}}}{OD_{\text{STANDARD}} - OD_{\text{BLANK}}} \times 300 \text{ (}\mu\text{g/dL)}$$

OD_{SAMPLE}, OD_{BLANK} and OD_{STANDARD} are optical density values of the Sample, Blank and the 300 µg/dL Standard, respectively.

Conversions: 100 µg/dL Cu equals 15.5 µM, 0.0001% or 1 ppm.

MATERIALS REQUIRED, BUT NOT PROVIDED


Pipeting devices and accessories. For 96-well plate assays: clear flat-bottom 96-well plates and plate reader. For cuvet assays: spectrophotometer and cuvetts for measuring OD at 356-362nm.

GENERAL CONSIDERATIONS

For scarce samples (e.g. mice serum or plasma), mix sample with dH₂O to a total of 100 µL, e.g. 50 µL serum + 50 µL dH₂O. Multiply the results by the dilution factor (2 fold).

EXAMPLES

Human serum, rat plasma, rat serum, and bovine serum were assayed in duplicate using the 96-well plate assay protocol. The copper concentrations were 97 ± 1, 104 ± 1, 101 ± 2, 78 ± 1 µg/dL, respectively.



Standard Curve in 96-well plate assay

PUBLICATIONS

1. Hendricks, M. R., et al (2021). Extracellular vesicles promote transkingdom nutrient transfer during viral-bacterial co-infection. Cell Reports, 34(4), 108672.

2. Matte, J. J., & Audet, I. (2020). Maternal perinatal transfer of vitamins and trace elements to piglets. Animal: An International Journal of Animal Bioscience, 14(1), 31-38.

3. Braunschmid, V., et al. (2020). A fungal ascorbate oxidase with unexpected laccase activity. International Journal of Molecular Sciences, 21(16).

